

Technical Specifications Compliance sheet for X-Ray Copper tube.

SI.No	Description of Item	Technical Specifications	CMTI Requirements	Compliance
01.	X- Ray Tube	2200 W, Copper (Cu), 0.4mm*12mm	<p>1.X-Ray tube should be compatible for existing Bruker D8 Advance X-Ray Diffractometer installed at CMTI. Serial No. of Equipment: 205044.</p> <p>2. Item should be installed & commissioned at CMTI.</p> <p>3. Copper X-Ray tube life should be 10,000 Hours</p> <p>4. Intensity of copper X-ray tube will be verified with Corundum Standard available at CMTI</p>	